


*High-throughput, low-volume,
multianalyte quantification of plasma
metabolites related to one-carbon
metabolism using HPLC-MS/MS*

**Øivind Midttun, Gry Kvalheim & Per
Magne Ueland**

**Analytical and Bioanalytical
Chemistry**

ISSN 1618-2642
Volume 405
Number 6

Anal Bioanal Chem (2013)
405:2009-2017
DOI 10.1007/s00216-012-6602-6



Your article is protected by copyright and all rights are held exclusively by Springer-Verlag Berlin Heidelberg. This e-offprint is for personal use only and shall not be self-archived in electronic repositories. If you wish to self-archive your work, please use the accepted author's version for posting to your own website or your institution's repository. You may further deposit the accepted author's version on a funder's repository at a funder's request, provided it is not made publicly available until 12 months after publication.

High-throughput, low-volume, multianalyte quantification of plasma metabolites related to one-carbon metabolism using HPLC-MS/MS

Øivind Midttun · Gry Kvalheim · Per Magne Ueland

Received: 30 August 2012 / Revised: 20 November 2012 / Accepted: 23 November 2012 / Published online: 13 December 2012
© Springer-Verlag Berlin Heidelberg 2012

Abstract Risk of chronic diseases, like cardiovascular disease and cancer, has been associated with biomarkers related to one-carbon metabolism, which comprises a metabolic network of cross-talking pathways. To address this complexity in epidemiological studies, we have established an isotope dilution HPLC-MS/MS method for quantification of 12 biomarkers and metabolites. All sample handling is performed by a robotic workstation. The assay uses 45 μ L of plasma, and sample treatment consists of protein precipitation by trichloroacetic acid. The analytes were separated on a Fortis Phenyl column using an isocratic mobile phase that contained water, methanol and acetic acid. Methionine, methionine sulfoxide, choline, betaine, dimethylglycine, arginine, asymmetric dimethylarginine, symmetric dimethylarginine, homoarginine, creatinine, cystathionine and trimethyllysine all showed limits of detection well below the 5th percentile of plasma distributions in healthy humans, coefficients of variation were in the range 2.2–12.3 %, and recoveries were 80–131 %. Simple sample processing, low-volume consumption, multiplexing and high capacity/short run time of this method make it suitable for large-scale metabolic profiling of precious biobank samples.

Keywords One-carbon metabolism · HPLC-MS/MS · High throughput · Epidemiology

Abbreviations

ADMA	Asymmetric dimethylarginine
BHMT	Betaine–homocysteine methyltransferase
CV	Coefficient of variation
DMG	Dimethylglycine
ESI	Electrospray ionization
EDTA	Ethylenediaminetetraacetic acid
HPLC-MS/MS	High-performance liquid chromatography tandem mass spectrometry
HSID	Hot source-induced desolvation
LOD	Limit of detection
MeOH	Methanol
MetSO ₄	Methionine sulfoxide
MRM	Multiple reaction mode
NO	Nitric oxide
NOS	Nitric oxide synthase
OCM	One-carbon metabolism
Q	Quadropole
S/N	Signal-to-noise ratio
SDMA	Symmetric dimethylarginine
TCA	Trichloroacetic acid
TML	Trimethyllysine

Electronic supplementary material The online version of this article (doi:10.1007/s00216-012-6602-6) contains supplementary material, which is available to authorized users.

Ø. Midttun (✉)

Bevital AS, Laboratory Building, 9th floor,
5021 Bergen, Norway
e-mail: nkjbm@uib.no

G. Kvalheim · P. M. Ueland

Section for Pharmacology, Institute of Medicine,
University of Bergen, 5021 Bergen, Norway

P. M. Ueland

Laboratory of Clinical Biochemistry, Haukeland University
Hospital, Bergen 5021, Norway

Introduction

Biomarkers related to one-carbon metabolism (OCM) have been associated with risk of chronic diseases, including cardiovascular disease [1] and cancer [2–4]. Methionine, a key component in OCM, is a precursor of the universal methyl donor *S*-adenosylmethionine [5]; it is obtained through diet or formed by remethylation of homocysteine. Alternatively, homocysteine is directed to cystathionine through the transsulfuration pathway. Homocysteine

remethylation is catalyzed by ubiquitous methionine synthase or (in the liver and kidney) by betaine–homocysteine methyltransferase (BHMT) requiring betaine as methyl donor [5]. Betaine is formed by oxidation of choline and is demethylated to dimethylglycine (DMG) during the BHMT reaction [2]. OCM also has ramifications to asymmetric dimethylarginine (ADMA), symmetric dimethylarginine (SDMA), creatinine and trimethyllysine (TML) since the syntheses of these metabolites all involve *S*-adenosylmethionine-dependent transmethylation [6–8].

Several metabolites of the outlined network have been assigned important roles in normal physiology, in the pathogenesis of human diseases or serve as useful biomarkers of disease risk. Choline and choline derivatives are components of structural lipoproteins, blood and membrane lipids. Choline is a precursor of the neurotransmitter acetylcholine, and its perinatal availability is important for neurodevelopment in rodents; betaine functions as an intracellular osmoregulator [2]. Recently, plasma choline and betaine have been shown to have opposite relations with key components of the metabolic syndrome [9]. Arginine [10], but also homoarginine [10] is a substrate of the enzyme nitric oxide synthase (NOS) forming nitric oxide (NO), a regulator of endothelial function; ADMA is an inhibitor and potential regulator of NOS activity, whereas SDMA may decrease arginine availability for NO synthesis [11]. Interference with NO synthesis may partly explain the observed associations of ADMA [12], SDMA [13–15] and homoarginine [16] with cardiovascular risk. In addition, SDMA is a marker of renal function, and may even outperform creatinine in this respect [17].

Most evidence on metabolite and disease risk referred to above has been obtained from epidemiological or clinical studies based on targeted plasma measurement of one or a few structurally related analytes. Methods used measure methionine and its oxidation product methionine sulfoxide (MetSO₄) by GC-MS [18], choline, betaine and DMG [19, 20] by LC-MS/MS or various combinations of arginine, ADMA, SDMA and homoarginine using either HPLC coupled with fluorescence detection [21–27] or mass spectrometry-based methods [28–31].

To save precious biobank material, we have established complementary platforms/methods which each simultaneously measures a panel of biomarkers or metabolites in a low volume of plasma. These analytes have been selected based on function and the actual biochemical pathway, rather than chemical class, i.e. quantitative metabolic profiling [32]. The present report describes a high-throughput, low-volume, multiplexing assay for profiling of 12 metabolites (methionine, MetSO₄, choline, betaine, dimethylglycine, arginine, ADMA, SDMA, homoarginine, creatinine, cystathionine and trimethyllysine) related to OCM.

Materials and methods

Chemicals

²H₄-Cystathionine (purity >95 %) was from C/D/N isotopes. Cystathionine (>95 %) and creatinine (>99 %) were purchased from Fluka. ²H₇-Arginine (98 %), ²H₇-ADMA (98 %), ¹³C₇¹⁵N₄-homoarginine (>98 %) and ²H₄-methionine (>99 %) were from Cambridge Isotope Laboratories. Arginine (98 %), MetSO₄ (>99 %), ²H₆-DMG (99 %), DMG (99 %), homoarginine (99.5 %), betaine (>99 %), methionine (>99 %), SDMA (>99 %), ADMA (98 %), ²H₉-choline (98 %), choline (>99 %), ²H₃-creatinine (98 %) and TML (>97 %) were from Sigma-Aldrich. ²H₂-SDMA (purity not given by producer) and ⁹H₂-TML (>95 %) were from PharmaAgra Labs. ²H₉-betaine (>99 %) was from Isotec.

Individual stock solutions of all analytes and internal standards were prepared in water and stored at –80 °C. Unlabelled compounds were not detected in any of the isotope-labelled internal standards. Using mass spectrometry, we investigated if isotopologues other than ²H₂-labelled was present in ²H₂-SDMA, and none were found.

A fraction of a plasma was spiked with all analytes, the unspiked and spiked plasma were analysed and the analyte concentrations in the unspiked plasma were estimated by using the standard addition method. Thus, the concentrations in the unspiked plasma were calculated as 0.330 µmol/L cystathionine, 73.5 µmol/L arginine, 2.30 µmol/L MetSO₄, 4.46 µmol/L DMG, 1.95 µmol/L homoarginine, 37.1 µmol/L betaine, 0.540 µmol/L SDMA, 32.0 µmol/L methionine, 0.550 µmol/L ADMA, 11.2 µmol/L choline, 70.2 µmol/L creatinine and 0.740 µmol/L TML. The unspiked and spiked portions of this plasma were aliquoted and stored at –80 °C, and the unspiked version was used as calibrator in routine use of the assay.

Instruments

A series 1100 HPLC system (Agilent Technologies) equipped with thermostated autosampler and degasser was used for solvent delivery and sample introduction. The HPLC system was coupled to an API 3000 triple-quadrupole tandem mass spectrometer (AB Sciex) equipped with an electrospray ion source and fitted with a hot source-induced desolvation (HSID) from IONICS. A column switcher from Valco (type VMHA) was used to divert the flow to waste during the first 1.2 min of each run and between each sample injection. Analyst (Ver. 1.5.2; AB Sciex) was used for data acquisition and analysis.

Sample collection and processing

Ethylenediaminetetraacetic acid (EDTA) plasma was obtained by collecting blood into Vacutainer Tubes (Becton Dickinson)

giving a final EDTA concentration of 4 mmol/L. The samples were immediately put on ice and centrifuged (at 2,000×g for 10 min at 4 °C) within 60 min. The plasma was collected and processed further, or stored at −80 °C until use.

Sample processing was performed by a robotic workstation (MikrolabAT Plus) fitted with disposable tips with liquid detection. In 96-well microtiter plates, 45 µL of plasma was mixed with 45 µL of trichloroacetic acid (TCA, 6 %) containing the internal standards (1.50 µmol/L ²H₄-cystathionine, 30.0 µmol/L ²H₇-arginine, 5.00 µmol/L ²H₆-DMG, 1.25 µmol/L ¹³C₇¹⁵N₄-homoarginine, 20.0 µmol/L ²H₉-betaine, 70.0 µmol/L ²H₄-methionine, 2.00 µmol/L ²H₂-SDMA, 1.00 µmol/L ²H₇-ADMA, 10.0 µmol/L ²H₉-choline, 94.0 µmol/L ²H₃-creatinine, 1 µmol/L ⁹H₂-TML) were added and the solution mixed by the robot. The mixture was then centrifuged at 6,000×g for 5 min at 5 °C, and the supernatant was transferred to a new plate which was placed in the autosampler compartment.

HPLC-MS/MS

Processed samples were kept in an autosampler at +6 °C in subdued light. Samples of 10 µL of deproteinized plasma were injected onto a Fortis Phenyl column (150×4.6 mm;

particle size, 3 µm) from Fortis Technologies guarded by a Phenomenex Polar-RP SecurityGuard Cartridge (4×3.0 mm). The temperature in the column compartment was +20 °C.

The mobile phase consisted of 2.5 % methanol (MeOH) in 150 mmol/L acetic acid (pH=5.2). Chromatography was isocratic, with a flow rate of 1.2 mL/min and run time of 3.4 min, and samples were injected every 4.2 min.

The mass spectrometer was operated in the positive ESI mode, and analytes and internal standards were detected in MRM with unit resolution at quadropole 1 (Q1) and low resolution at Q3 (in order to increase the sensitivity for low-abundance analytes). Other common instrument parameters were nebulizer gas=7 psig, curtain gas=12 psig, collision-activated dissociation gas=3 psig, ion spray=5,000 V, ion source temperature=500 °C, declustering potential=23 V, focusing potential=30 V and HSID temperature=200 °C. For each compound and internal standard, the ion pair used consisted of the singly protonated molecular ion and its most abundant product ion. Table 1 lists the ion pairs, together with retention times and ion-pair specific instrument parameters for all analytes and internal standards. We used isotopologues as internal standards for all analytes when available (²H₆-DMG was used for MetSO₄), and quantitation was performed by dividing

Table 1 Retention times and instrument settings

Analyte	<i>t_r</i> , min	Transitions, <i>m/z</i>		EP, V	CE, V	CXP, V
		Precursor ion	Product ion			
⁴ H ₂ -Cystathionine	1.48	226.9	137.9	6	20	8
Cystathionine	1.48	222.9	133.9	6	20	8
⁷ H ₂ -Arginine	1.69	181.9	77.2	10	27	5.6
Arginine	1.70	174.9	70.2	10	27	5.6
MetSO ₄	1.75	166.1	73.9	10	20	5
⁶ H ₂ -DMG	1.81	110.1	64.1	10	20	10
DMG	1.82	104.1	58.1	10	20	10
¹³ C ₇ ¹⁵ N ₄ -Homoarginine	1.89	200.2	90.2	10	30	6
Homoarginine	1.90	189.2	84.1	10	30	6
⁹ H ₂ -Betaine	2.18	127.1	68.1	8	28	4
Betaine	2.20	118.1	59.1	8	28	4
⁴ H ₂ -Methionine	2.34	154.2	108.0	8	15	6
Methionine	2.35	150.2	104.0	8	15	6
SDMA	2.39	203.2	172.1	10	30	7
² H ₂ -SDMA	2.51	205.2	72.2	10	35	7
⁷ H ₂ -ADMA	2.60	210.2	77.2	10	24	7
ADMA	2.63	203.2	46.1	10	24	7
⁹ H ₂ -Choline	2.66	113.1	69.1	12	28	4
Choline	2.73	104.1	60.1	12	28	4
³ H ₂ -Creatinine	3.15	117.2	47.2	14	28	4
Creatinine	3.19	114.2	44.2	14	28	4
⁹ H ₂ -TML	2.05	198.1	84.2	10	30	6
TML	2.05	189.2	84.1	10	30	6

t_r, retention time, *EP* entrance potential, *CE* collision energy, *CXP* collision cell exit potential.

the analyte peak area by the area of the corresponding internal standard.

Method validation

Linearity was investigated in solutions of 4 % albumin in phosphate-buffered saline. Limit of detection (LOD) was determined from the linearity experiments as signal-to-noise ratio (S/N, determined using the script supplied with the Analyst software by Applied Biosystems) of 3.

Within-day and between-day coefficients of variation (CV) were determined by analysing a plasma pooled from three healthy individuals at three concentrations (low=endogenous, medium and high spiking concentrations) of each analyte. For the within-day experiments, 19 replicates of the same samples were analysed on the same day, while for the between-day experiments the same samples were analysed twice on 19 different days. From the within-day CV experiments, recoveries were determined. The spiking levels for the CV and recovery experiments are given in Table 2.

Matrix effects were studied by analysing plasma from ten persons and water. Both plasma and water were spiked after sample treatment with the high concentrations used in the CV and recovery experiments (Table 2), and calculated as

Matrix effect

$$= \frac{[\text{analyte peak area}]_{\text{spiked}} - [\text{analyte peak area}]_{\text{endogenous}}}{[\text{analyte peak area}]_{\text{water}}} \times 100\%$$

and also by replacing the analyte peak areas in the formula by the analyte peak area divided by the internal standard

peak area. The relative matrix effect, which describes the spread of matrix effects, was calculated as the CV of the matrix effects of the ten plasma samples. $^2\text{H}_6$ -DMG was selected as internal standard for MetSO₄ based on the method validation results and similar retention times.

In order to establish the analyte concentration in healthy subjects, we analysed plasma concentrations in 171 healthy Norwegian adult blood donors [39.8 % female, average age 44 (range, 21–68)years].

Results and discussion

The presented assay analyses 12 plasma biomarkers (Figure S1, Electronic supplementary material) related to OCM. Sample treatment is very quick and simple, and involves protein precipitation only. We exploited the specificity and sensitivity of combining HPLC with tandem mass spectrometry to develop an assay which requires only 45 µL of plasma and has a chromatographic run time of 3.8 min.

Sample treatment, stability of analytes and internal standards

Sample preparation consisted of precipitation of proteins by mixing the plasma with TCA (performed by a robotic workstation) followed by centrifugation, a simple and quick procedure, enabling high-throughput analyses. The final concentration of TCA was such that more than 99.5 % of the plasma proteins removed from the plasma [33], giving a clean supernatant ensuring long column lifetimes and reducing the need for ion-source cleaning. On average, columns were replaced every 1,000 injections.

Table 2 Analytical recovery of the assay

Analyte	Concentration (µM) ^a			Recovery ^b , %	
	Endogenous	Endogenous+added		Added	
		Low	Medium	High	Medium
Cystathionine	0.22	1.13	2.13	92	96
Arginine	73.0	137.9	206.3	89	90
MetSO ₄	1.80	3.31	4.98	98	106
DMG	3.9	7.7	11.8	125	131
Homoarginine	1.73	4.62	7.37	118	114
Betaine	38	61	83	91	90
Methionine	18.9	50.4	83.0	105	107
SDMA	0.47	1.32	2.16	84	84
ADMA	0.45	1.72	3.02	102	103
Choline	8.4	16.6	24.8	80	82
Creatinine	79.3	141.1	205.5	91	91
TML	0.52	1.52	2.51	66	66

^an=19 for all concentrations

^bData are given as mean values of the 19 runs (same as for the within-day CV experiments)

Published information about analyte stability was used as a guide in establishing sample treatment and storage conditions. Plasma methionine is stable during short-term storage at room temperature, but may be oxidized to MetSO₄ when stored for years at -25 °C [34]. Thus, the sum of methionine plus MetSO₄ represents plasma total methionine [34]. Choline concentrations in EDTA plasma increase during storage at room temperature [34, 35], but are stable for at least 72 h at 0 °C and for years at -80 °C [19]. Betaine and DMG are stable in EDTA plasma at 0 °C for at least 72 h [19, 34] and for at least 29 years at -25 °C [34]. Arginine, ADMA and SDMA are stable in EDTA plasma for several years at -20 °C [21] and for at least 8 days at room temperature [21, 27, 34]. Homoarginine is stable in EDTA plasma for at least 8 days at room temperature [27]. Cystathionine is stable in plasma at room temperature for up to 8 days, but degrades if stored for several years at -25 °C [34]. Creatinine is stable [34].

No changes in signal intensities for analytes or internal standards were observed for treated plasma samples stored for up to 48 h in the autosampler compartment in subdued light at +6 °C, demonstrating that the analytes and internal standards were stable under these conditions.

HPLC-MS/MS

Selection of mobile phase and column was optimized to give the best combination of signal intensities and method performance for the included analytes. Optimization of the amount of MeOH in the mobile phase ensured adequate peak shapes for all analytes, and all analytes eluted within

3.2 min (Table 1 and Fig. 1). The high concentration of acetic acid in the mobile phase provides good ionization of the analytes [36].

Selection of ion-pairs for analytes and internal standards were based on peak intensities and S/N ratios, and the selected ion-pairs are listed in Table 1. We avoided using mass transitions that corresponds to loss of water or ammonia to avoid interference from coeluting isobaric compounds that may also lose such fragments. The combined use of chromatographic separation and tandem mass spectrometry thus provided chromatograms that were usually without interfering peaks (Fig. 1). Homoarginine and TML, which have identical masses and primary product ions, were separated by carefully optimizing the chromatographic conditions. TML eluted between two interfering peaks, but sufficient chromatographic resolution was obtained to give adequate assay performance (see below) also for this analyte. Acquisition parameters were optimized for each compound, and are listed in Table 1.

Matrix effects

Matrix effects were investigated in plasma from ten persons and calculated from analyte peak areas only and also for the analyte/ISTD peak area ratios (Table 3). The range of matrix effects was 32–99 %, and relative matrix effect was below 6 % for all analytes except MetSO₄ (12 %) and DMG (11 %). After adjustment for internal standards, the matrix effects for all compounds (except MetSO₄) were in the range 90–104 %, and the relative matrix effects were reduced for 7 of the 11 analytes. For MetSO₄, the matrix effect

Fig. 1 Normalized HPLC-MS/MS chromatogram of the calibrator plasma (spiked with MetSO₄) containing 1.50 μM ²H₄-cystathionine, 0.33 μM cystathionine, 30.0 μM ²H₇-arginine, 73.5 μM arginine, 6.59 μM MetSO₄, 5.00 μM ²H₆-DMG, 4.5 μM DMG, 1.25 μM ¹³C₇¹⁵N₄-homoarginine, 1.95 μM homoarginine, 20.0 μM ²H₉-betaine, 37 μM betaine, 70.0 μM ²H₄-methionine, 32.0 μM methionine, 2.00 μM ²H₂-SDMA, 0.54 μM SDMA, 1.00 μM ²H₇-ADMA, 0.55 μM ADMA, 10.0 μM ²H₉-choline, 11.2 μM choline, 94.0 μM ²H₃-creatinine and 70.2 μM creatinine




Table 3 Matrix effects

Analyte	Based on analyte area	Based on analyte/ISTD area ratio
Cystathionine	39 (5)	101 (2)
Arginine	57 (4)	92 (3)
MetSO ₄	32 (11)	45 (10)
DMG	72 (12)	99 (12)
Homoarginine	96 (5)	93 (7)
Betaine	71 (5)	96 (3)
Methionine	94 (4)	99 (2)
SDMA	88 (4)	90 (7)
ADMA	99 (6)	104 (7)
Choline	78 (6)	99 (4)
Creatinine	75 (5)	92 (3)
TML	69 (7)	89 (6)

The matrix effects are based on ten different plasma samples. All values (percentage) are given as mean (CV)

ISTD internal standard

was generally somewhat reduced, while the relative matrix effect was generally not improved, by adjustment with different internal standards (data shown using only ²H₆-DMG as internal standard, Table 3).

CV, recovery, LOD and linearity

CVs were 2.2–12.3 % (Table 4), and recoveries were in the range 66–131 % (Table 2). LODs (Table 5) were all below the 5th percentile of the concentrations found in 171 healthy subjects (Table 6). The linear ranges (Table 5) spanned the endogenous concentrations observed in plasma/serum from healthy humans (Table 6).

Table 4 Imprecision of the assay

Analyte	Within-day CV (<i>N</i> =19) ^a , %			Between-day CV (<i>N</i> =19), %		
	Low	Medium	High	Low	Medium	High
Cystathionine	12.0	8.0	5.9	10.6	10.3	7.0
Arginine	5.6	4.5	3.6	2.6	5.5	2.2
MetSO ₄	12.3	8.8	9.0	7.9	7.6	7.6
DMG	11.7	8.9	6.7	7.7	9.4	5.0
Homoarginine	6.9	5.9	6.4	5.5	8.0	5.1
Betaine	7.2	7.3	5.5	8.3	9.5	7.8
Methionine	6.5	5.4	4.3	5.1	7.0	3.2
SDMA	7.7	7.9	8.6	7.7	9.1	6.4
ADMA	7.0	5.2	6.6	7.8	8.5	6.5
Choline	5.4	7.6	5.9	5.4	6.8	3.8
Creatinine	4.6	4.6	3.4	2.2	5.8	2.7
TML	7.3	6.0	5.6	8.7	9.8	12.2

^aData from the recovery experiments

Method validation characteristics were generally similar to those reported for other methods for arginine [27, 30], MetSO₄ [18], DMG [19, 20], homoarginine [27], betaine [19, 20], SDMA [27, 30], methionine [18], ADMA [27, 30], choline [19, 20] and creatinine [37].

Recoveries that deviate significantly from 100 % are common [38], and we observed high recovery for DMG (125–131 %) and low recovery for TML (66 %). Such marked deviations from 100 % may reflect uneven distribution of analytes between the supernatant and pellet, and may affect the accuracy of an assay.

Assays using much larger sample volumes have shown better method characteristics for cystathionine and methionine [39, 40] than the current method. Better method performance has also been demonstrated for cystathionine by using a mass spectrometer with higher sensitivity combined with a higher concentration of acetic acid in the mobile phase [36], and for homoarginine by a more time-consuming, single analyte assay [41]. For methods including different combinations of arginine, homoarginine, SDMA and ADMA, similar or better validation characteristics have been published [22–31]; however, all but one of these methods require higher sample volume (100–1000 µL) [22–29, 31] and/or longer chromatographic separation time (up to 52 min) [22–28, 31] than the current assay. The combination of low volume consumption, multiplexing, high throughput and good method performance makes the presented assay useful for metabolic profiling in large-scale studies.

Plasma concentrations in healthy subjects

Plasma concentrations were investigated in 171 nonfasting healthy adult blood donors (Table 6). Median concentrations

Table 5 Limit of detection and linearity

Analyte	Linear range (μM)	LOD (μM)	Regression parameters		
			Slope	Intercept	r^2
Cystathionine	0.025–40	0.025	0.5895	0.1247	0.999
Arginine	0.25–200	0.25	0.0403	0.0771	0.999
MetSO ₄	0.14–229	0.14	0.0144	0.0225	0.987
DMG	1.25–80	1.25	0.0227	-0.0020	0.999
Homoarginine	0.10–40	0.10	1.0123	0.0143	0.998
Betaine	0.50–400	0.50	0.0174	0.0738	0.998
Methionine	1.0–100	1.0	0.0102	0.0009	0.999
SDMA	0.080–2.5	0.080	0.7875	0.0504	0.995
ADMA	0.080–5.0	0.080	0.5084	-0.0018	0.985
Choline	0.50–100	0.50	0.0166	0.0109	0.997
Creatinine	0.25–800	0.25	0.00065	0.0423	0.998
TML	0.20–20	0.20	0.3987	0.2581	0.983

Linearity performed in PBS

LOD limit of detection (defined as $S/N > 3$), LLQ lower limit of quantification ($LLQ = 2 \times LOD$)

($\mu\text{mol/L}$) were 0.26 for cystathionine, 115.0 for arginine, 3.65 for MetSO₄, 4.0 for DMG, 1.93 for homoarginine, 39 for betaine, 0.49 for SDMA, 30.4 for methionine, 0.50 for ADMA, 11.3 for choline, 79.2 for creatinine and 0.68 for TML. These concentrations were similar to those reported previously for cystathionine [42, 43], arginine [26, 27, 29, 30], MetSO₄ [18], homoarginine [24, 27, 41], betaine [43, 44], SDMA [24, 27, 29, 30], methionine [18, 43, 45], ADMA [24, 27, 29, 30], choline [44], creatinine [42, 44] and TML [46] (Table 6).

Defining the lower limit of quantification (LLQ) as two times LOD (Table 5), Table 6 shows that the 5th percentile of all analytes except DMG and TML are well above LLQ. For quantification of DMG and TML a lower precision would be expected at the lower end of normal plasma concentrations.

Plasma concentrations of MetSO₄ were calculated by dividing the area of MetSO₄ by the areas of several of the available internal standards. Comparing the resulting concentrations by Passing–Bablok regression showed that there were essentially no differences between the concentrations obtained by using different internal standards (not shown).

Conclusions

We have developed a low-volume, high-throughput HPLC-MS/MS assay that determines 12 plasma biomarkers related to OCM. The assay is presently being used for metabolic profiling in large-scale epidemiologic studies of chronic diseases.

Table 6 Analyte concentrations (μM) in plasma or serum

Analyte	This work ^a		Literature	
	Median	5–95 percentile	Median or mean	Reference
Cystathionine	0.26	(0.13–0.64)	0.19, 0.11	42, 43
Arginine	115.0	(81.3–163.5)	72.6, 94.2, 85, 85	26, 27, 29, 30
MetSO ₄	3.65	(2.18–6.59)	4	18
DMG	4.0	(2.4–6.4)	2.5, 3.1	43, 44
Homoarginine	1.93	(1.17–3.31)	2.15, 2.5, 2.5	24, 27, 41
Betaine	39	(25–58)	35, 30.3	43, 44
Methionine	30.4	(21.0–43.5)	28.3, 20, 26.5	18, 43, 45
SDMA	0.49	(0.38–0.65)	0.40, 0.41, 0.6, 0.5	24, 27, 29, 30
ADMA	0.50	(0.39–0.65)	0.45, 0.48, 0.6, 0.5	24, 27, 29, 30
Choline	11.3	(8.6–15.0)	7.8	44
Creatinine	79.2	(58.9–99.3)	68.9, 74	42, 44
TML	0.68	(0.40–1.31)	0.33	46

^aPlasma from 171 nonfasting individuals

Acknowledgments This study was funded by the non-profit “Foundation to Promote Research into Functional Vitamin B12 Deficiency”, the Western Norway Regional Health Authority and the Norwegian Cancer Society. We thank Saleh Rayyan and Marit Krokeide for technical assistance.

References

- Smulders YM, Blom HJ (2011) The homocysteine controversy. *J Inherit Metab Dis* 34(1):93–99. doi:10.1007/s10545-010-9151-1
- Ueland PM (2011) Choline and betaine in health and disease. *J Inherit Metab Dis* 34(1):3–15. doi:10.1007/s10545-010-9088-4
- Duthie SJ (2011) Folate and cancer: how DNA damage, repair and methylation impact on colon carcinogenesis. *J Inherit Metab Dis* 34(1):101–109. doi:10.1007/s10545-010-9128-0
- Mangoni AA (2009) The emerging role of symmetric dimethylarginine in vascular disease. *Adv Clin Chem* 48:73–94
- Finkelstein JD (1990) Methionine metabolism in mammals. *J Nutr Biochem* 1:228–237
- Teerlink T (2005) ADMA metabolism and clearance. *Vasc Med* 10 (Suppl 1):S73–S81
- Tran CT, Leiper JM, Vallance P (2003) The DDAH/ADMA/NOS pathway. *Atheroscler Suppl* 4(4):33–40
- Brosnan JT, Brosnan ME (2007) Creatine: endogenous metabolite, dietary, and therapeutic supplement. *Annu Rev Nutr* 27:241–261. doi:10.1146/annurev.nutr.27.061406.093621
- Konstantinova SV, Tell GS, Vollset SE, Nygard O, Bleie O, Ueland PM (2008) Divergent associations of plasma choline and betaine with components of metabolic syndrome in middle age and elderly men and women. *J Nutr* 138(5):914–920
- Moali C, Boucher JL, Sari MA, Stuehr DJ, Mansuy D (1998) Substrate specificity of NO synthases: detailed comparison of L-arginine, homo-L-arginine, their N omega-hydroxy derivatives, and N omega-hydroxynor-L-arginine. *Biochemistry* 37(29):10453–10460. doi:10.1021/bi980742t
- Moncada S, Higgs A (1993) The L-arginine-nitric oxide pathway. *N Engl J Med* 329(27):2002–2012. doi:10.1056/NEJM199312303292706
- Cooke JP (2004) Asymmetrical dimethylarginine: the Uber marker. *Circulation* 109(15):1813–1818. doi:10.1161/01.CIR.0000126823.07732.D5
- Bode-Boger SM, Scalera F, Kielstein JT, Martens-Lobenhoffer J, Breithardt G, Fobker M, Reinecke H (2006) Symmetrical dimethylarginine: a new combined parameter for renal function and extent of coronary artery disease. *J Am Soc Nephrol* 17(4):1128–1134. doi:10.1681/ASN.2005101119
- Fliser D, Kronenberg F, Kielstein JT, Morath C, Bode-Boger SM, Haller H, Ritz E (2005) Asymmetric dimethylarginine and progression of chronic kidney disease: the mild to moderate kidney disease study. *J Am Soc Nephrol* 16(8):2456–2461. doi:10.1681/ASN.2005020179
- Meinitzer A, Kielstein JT, Pilz S, Drechsler C, Ritz E, Boehm BO, Winkelmann BR, Marz W (2011) Symmetrical and asymmetrical dimethylarginine as predictors for mortality in patients referred for coronary angiography: the Ludwigshafen Risk and Cardiovascular Health study. *Clin Chem* 57(1):112–121. doi:10.1373/clinchem.2010.150854
- Marz W, Meinitzer A, Drechsler C, Pilz S, Krane V, Kleber ME, Fischer J, Winkelmann BR, Bohm BO, Ritz E, Wanner C (2010) Homoarginine, cardiovascular risk, and mortality. *Circulation* 122(10):967–975. doi:10.1161/CIRCULATIONAHA.109.908988
- Kielstein JT, Veldink H, Martens-Lobenhoffer J, Haller H, Burg M, Lorenzen JM, Lichtinghagen R, Bode-Boger SM, Kliem V (2011) SDMA is an early marker of change in GFR after living-related kidney donation. *Nephrol Dial Transplant* 26(1):324–328. doi:10.1093/ndt/gfq395
- Mashima R, Nakanishi-Ueda T, Yamamoto Y (2003) Simultaneous determination of methionine sulfoxide and methionine in blood plasma using gas chromatography–mass spectrometry. *Anal Biochem* 313(1):28–33
- Holm PI, Ueland PM, Kvalheim G, Lien EA (2003) Determination of choline, betaine, and dimethylglycine in plasma by a high-throughput method based on normal-phase chromatography–tandem mass spectrometry. *Clin Chem* 49(2):286–294
- Kirsch SH, Herrmann W, Rabagny Y, Obeid R (2010) Quantification of acetylcholine, choline, betaine, and dimethylglycine in human plasma and urine using stable-isotope dilution ultra performance liquid chromatography–tandem mass spectrometry. *J Chromatogr B Analyt Technol Biomed Life Sci* 878(32):3338–3344. doi:10.1016/j.jchromb.2010.10.016
- Teerlink T (2007) HPLC analysis of ADMA and other methylated L-arginine analogs in biological fluids. *J Chromatogr B Analyt Technol Biomed Life Sci* 851(1–2):21–29. doi:10.1016/j.jchromb.2006.07.024
- Blackwell S, O'Reilly DS, Talwar D (2007) Biological variation of asymmetric dimethylarginine and related arginine metabolites and analytical performance goals for their measurement in human plasma. *Eur J Clin Invest* 37(5):364–371. doi:10.1111/j.1365-2362.2007.01798.x
- Blackwell S, O'Reilly DS, Talwar DK (2009) HPLC analysis of asymmetric dimethylarginine (ADMA) and related arginine metabolites in human plasma using a novel non-endogenous internal standard. *Clin Chim Acta* 401(1–2):14–19. doi:10.1016/j.cca.2008.10.032
- Jones CE, Darcy CJ, Woodberry T, Anstey NM, McNeil YR (2010) HPLC analysis of asymmetric dimethylarginine, symmetric dimethylarginine, homoarginine and arginine in small plasma volumes using a Gemini-NX column at high pH. *J Chromatogr B Analyt Technol Biomed Life Sci* 878(1):8–12. doi:10.1016/j.jchromb.2009.10.035
- Marra M, Bonfigli AR, Testa R, Testa I, Gambini A, Coppa G (2003) High-performance liquid chromatographic assay of asymmetric dimethylarginine, symmetric dimethylarginine, and arginine in human plasma by derivatization with naphthalene-2,3-dicarboxaldehyde. *Anal Biochem* 318(1):13–17
- Teerlink T, Nijveldt RJ, de Jong S, van Leeuwen PA (2002) Determination of arginine, asymmetric dimethylarginine, and symmetric dimethylarginine in human plasma and other biological samples by high-performance liquid chromatography. *Anal Biochem* 303(2):131–137. doi:10.1006/abio.2001.5575S0003269701955757
- Meinitzer A, Puchinger M, Winklhofer-Roob BM, Rock E, Ribalta J, Roob JM, Sundl I, Halwachs-Baumann G, Marz W (2007) Reference values for plasma concentrations of asymmetrical dimethylarginine (ADMA) and other arginine metabolites in men after validation of a chromatographic method. *Clin Chim Acta* 384(1–2):141–148. doi:10.1016/j.cca.2007.07.006
- Weaving G, Rocks BF, Bailey MP, Titheradge MA (2008) Arginine and methylated arginines in human plasma and urine measured by tandem mass spectrometry without the need for chromatography or sample derivatization. *J Chromatogr B Analyt Technol Biomed Life Sci* 874(1–2):27–32. doi:10.1016/j.jchromb.2008.08.016
- Huang LF, Guo FQ, Liang YZ, Li BY, Cheng BM (2004) Simultaneous determination of L-arginine and its mono- and dimethylated metabolites in human plasma by high-performance liquid chromatography–mass spectrometry. *Anal Bioanal Chem* 380(4):643–649. doi:10.1007/s00216-004-2759-y
- El-Khoury JM, Bunch DR, Reineks E, Jackson R, Steinle R, Wang S (2012) A simple and fast liquid chromatography–tandem mass spectrometry method for measurement of underivatized L-

- arginine, symmetric dimethylarginine, and asymmetric dimethylarginine and establishment of the reference ranges. *Anal Bioanal Chem* 402(2):771–779. doi:10.1007/s00216-011-5462-9
31. Davids M, Swieringa E, Palm F, Smith DE, Smulders YM, Scheffer PG, Blom HJ, Teerlink T (2012) Simultaneous determination of asymmetric and symmetric dimethylarginine, l-monomethylarginine, l-arginine, and l-homoarginine in biological samples using stable isotope dilution liquid chromatography tandem mass spectrometry. *J Chromatogr B Analyt Technol Biomed Life Sci* 900:38–47. doi:10.1016/j.jchromb.2012.05.025
 32. Ueland PM, Midttun O, Windelberg A, Svardal A, Skalevik R, Hustad S (2007) Quantitative profiling of folate and one-carbon metabolism in large-scale epidemiological studies by mass spectrometry. *Clin Chem Lab Med* 45(12):1737–1745. doi:10.1515/CCLM.2007.339
 33. Blanchard J (1981) Evaluation of the relative efficacy of various techniques for deproteinizing plasma samples prior to high-performance liquid chromatographic analysis. *J Chromatogr* 226(2):455–460
 34. Hustad S, Eussen S, Midttun O, Ulvik A, van de Kant PM, Morkrid L, Gislefoss R, Ueland PM (2012) Kinetic modeling of storage effects on biomarkers related to B vitamin status and one-carbon metabolism. *Clin Chem* 58(2):402–410. doi:10.1373/clinchem.2011.174490
 35. Yue B, Pattison E, Roberts WL, Rockwood AL, Danne O, Lueders C, Mockel M (2008) Choline in whole blood and plasma: sample preparation and stability. *Clin Chem* 54(3):590–593. doi:10.1373/clinchem.2007.094201
 36. Midttun O, Hustad S, Ueland PM (2009) Quantitative profiling of biomarkers related to B-vitamin status, tryptophan metabolism and inflammation in human plasma by liquid chromatography/tandem mass spectrometry. *Rapid Commun Mass Spectrom* 23:1371–1379
 37. Tsikas D, Wolf A, Mitschke A, Gutzki FM, Will W, Bader M (2010) GC-MS determination of creatinine in human biological fluids as pentafluorobenzyl derivative in clinical studies and biomonitoring: inter-laboratory comparison in urine with Jaffe, HPLC and enzymatic assays. *J Chromatogr B Analyt Technol Biomed Life Sci* 878(27):2582–2592. doi:10.1016/j.jchromb.2010.04.025
 38. Marchi I, Viette V, Badoud F, Fathi M, Saugy M, Rudaz S, Veuthey JL (2010) Characterization and classification of matrix effects in biological samples analyses. *J Chromatogr A* 1217(25):4071–4078. doi:10.1016/j.chroma.2009.08.061
 39. Melnyk S, Pogribna M, Pogribny I, Hine RJ, James SJ (1999) A new HPLC method for the simultaneous determination of oxidized and reduced plasma amino thiols using coulometric electrochemical detection. *J Nutr Biochem* 10(8):490–497
 40. Sigit JL, Hages M, Brensing KA, Frotscher U, Pietrzik K, von Bergmann K, Lutjohann D (2001) Total plasma homocysteine and related amino acids in end-stage renal disease (ESRD) patients measured by gas chromatography–mass spectrometry—comparison with the Abbott IMx homocysteine assay and the HPLC method. *Clin Chem Lab Med* 39(8):681–690
 41. Atzler D, Mieth M, Maas R, Bøger RH, Schwedhelm E (2011) Stable isotope dilution assay for liquid chromatography–tandem mass spectrometric determination of l-homoarginine in human plasma. *J Chromatogr B* 879:2294–2298
 42. Midttun O, Hustad S, Schneede J, Grotmol T, Hoff G, Ueland P (2007) Plasma vitamin B6 forms and their relations to transsulfuration metabolites in a large-scale population-based study. *Am J Clin Nutr* 86:131–138
 43. Allen RH, Stabler SP, Savage DG, Lindenbaum J (1993) Metabolic abnormalities in cobalamin (vitamin B12) and folate deficiency. *FASEB J* 7(14):1344–1353
 44. Holm PI, Ueland PM, Vollset SE, Midttun O, Blom HJ, Keijzer MB, den Heijer M (2005) Betaine and folate status as cooperative determinants of plasma homocysteine in humans. *Arterioscler Thromb Vasc Biol* 25(2):379–385
 45. Windelberg A, Arseth O, Kvalheim G, Ueland PM (2005) Automated assay for the determination of methylmalonic acid, total homocysteine, and related amino acids in human serum or plasma by means of methylchloroformate derivatization and gas chromatography–mass spectrometry. *Clin Chem* 51(11):2103–2109
 46. Ringseis R, Hanisch N, Seliger G, Eder K (2010) Low availability of carnitine precursors as a possible reason for the diminished plasma carnitine concentrations in pregnant women. *BMC Pregnancy Childbirth* 10:17. doi:10.1186/1471-2393-10-17